

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information		
Status: Published		Document ID: 57011
DATE EFFECTIVE 06/23/2025	APPROVED BY Nuclear DNA Technical Leader	PAGE 1 OF 32

Appendix for PowerPlex® Fusion Stutter

Table 1. Locations without STRmix™ stutter modeling organized (a) by location and (b) by stutter type for STRmix™ and GeneMarker®. *There are no DBS filters in GeneMarker®. (X)-marked locations are NOT modeled by STRmix™. Note: All locations model back stutter and are not included in the tables.

(a) Stutter modeling by location

	D3S1358	D1S1656	D2S441	D10S1248	D13S317	Penta E
FS						
HBS						X
DBS*	X			X	X	X
	D16S539	D18S51	D2S1338	CSF1PO	Penta D	
FS				X	X	
HBS	X			X	X	
DBS*	X			X	X	
	TH01	vWA	D21S11	D7S820	D5S818	TPOX
FS	X					X
HBS		X	X		X	X
DBS*	X	X		X	X	X
	D8S1179	D12S391	D19S433	FGA	D22S1045	
FS						
HBS	X	X	X	X	X	
DBS*			X		X	

(b) Stutter modeling by stutter type

	FS	HBS	DBS*		FS	HBS	DBS*
D3S1358			X	TH01	X		X
D1S1656				vWA		X	X
D2S441				D21S11		X	
D10S1248			X	D7S820			X
D13S317			X	D5S818		X	X
Penta E		X	X	TPOX	X	X	X
D16S539		X	X	D8S1179		X	
D18S51				D12S391		X	
D2S1338				D19S433		X	X
CSF1PO	X	X	X	FGA		X	
Penta D	X	X	X	D22S1045		X	X

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information		
Status: Published		Document ID: 57011
DATE EFFECTIVE 06/23/2025	APPROVED BY Nuclear DNA Technical Leader	PAGE 2 OF 32

Table 2. Maximum allowable stutter ratios used by STRmix™ per stutter type modeled. A peak in stutter position will only be modeled by STRmix™ as stutter if its peak height ratio to the parent peak is below this threshold.

Stutter Type	Maximum Allowable Stutter Ratio
Back Stutter	0.3
Forward Stutter	0.2
Half Back Stutter	0.1
Double Back Stutter	0.1

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information		
Status: Published		Document ID: 57011
DATE EFFECTIVE 06/23/2025	APPROVED BY Nuclear DNA Technical Leader	PAGE 3 OF 32

Table 3. Methods for stutter modeling used by STRmix™ and GeneMarker® by location. For locations that use longest uninterrupted sequence (LUS) or Average, if value is not available, the allele regression line is then used. *There are no DBS filters in GeneMarker®.

Locus	Back Stutter	Forward Stutter	Half Back Stutter	Double Back Stutter*
D3S1358	Average	Average	Average	
D1S1656	LUS	Average	Allele Regression	Average
D2S441	Average	Average	Average	Average
D10S1248	Allele Regression	Average	Average	
D13S317	Allele Regression	Average	Average	
Penta E	Average	Average		
D16S539	Allele Regression	Average		
D18S51	Allele Regression	Average	Average	Average
D2S1338	Average	Average	Allele Regression	Allele Regression
CSF1PO	Allele Regression			
Penta D	Average			
TH01	LUS		Average	
vWA	Average	Average		
D21S11	Average	Average		Average
D7S820	Allele Regression	Average	Average	
D5S818	Allele Regression	Average		
TPOX	Average			
DYS391				
D8S1179	Average	Average		Average
D12S391	Allele Regression	Average		Allele Regression
D19S433	Average	Average		
FGA	Average	Average		Average
D22S1045	Allele Regression	Average		

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information		
Status: Published		Document ID: 57011
DATE EFFECTIVE 06/23/2025	APPROVED BY Nuclear DNA Technical Leader	PAGE 4 OF 32

Table 4. Allele-specific stutter filters used in GeneMarker® analysis at NYC OCME. Values highlighted in blue show allele-specific averages. LUS assignments and values based on LUS calculations are highlighted in orange (D1S1656 and TH01).

D3S1358				D1S1656				D2S441				D10S1248				D13S317				Penta E				
Allele	BS	FS	HBS	Allele	LUS	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS
8	3.15%	1.50%	2.34%	8	8	7.29%	6.12%	2.42%	8	8.01%	8.25%	2.25%	8	3.18%	3.68%	1.89%	5	1.47%	3.03%	3.34%	4	1.81%	7.13%	
9	3.94%	1.64%	2.37%	9	9	8.26%	6.09%	2.48%	9	7.95%	8.32%	2.43%	9	4.08%	3.84%	1.89%	6	2.20%	3.26%	3.40%	5	2.11%	1.00%	
10	4.74%	1.78%	2.41%	10	9.5	8.75%	6.06%	2.53%	9.1	3.58%	8.33%	2.44%	10	4.97%	4.00%	1.90%	7	2.94%	3.49%	3.45%	6	2.40%	7.17%	
11	5.54%	1.92%	2.44%	11	11	10.20%	12.44%	2.59%	10	5.99%	7.36%	0.77%	11	5.87%	4.17%	1.90%	7.1	3.02%	3.52%	3.46%	7	1.29%	7.18%	
12	6.33%	2.06%	2.47%	12	11.5	10.69%	3.91%	2.65%	10.1	7.89%	8.40%	2.62%	12	6.76%	7.02%	1.91%	8	3.68%	2.52%	3.15%	8	2.05%	7.20%	
13	7.13%	2.20%	2.50%	13	12	11.17%	8.86%	2.70%	11	8.02%	4.52%	1.21%	13	7.66%	2.69%	1.92%	8.1	3.76%	3.75%	3.51%	9	3.30%	7.22%	
14	7.57%	1.13%	2.54%	13.3	11	10.20%	5.96%	2.72%	11.3	1.97%	15.86%	2.83%	14	8.55%	5.87%	1.92%	9	4.42%	0.42%	2.96%	10	2.09%	1.76%	
14.3	8.17%	2.39%	2.55%	14	13.5	12.63%	5.24%	2.76%	12	11.11%	1.88%	2.95%	15	9.44%	4.00%	1.26%	10	5.16%	2.94%	3.04%	11	4.55%	7.25%	
15	8.48%	1.03%	1.82%	14.3	9	8.26%	5.93%	2.78%	12.3	7.78%	8.57%	3.00%	16	10.34%	6.25%	1.93%	11	5.90%	3.75%	3.51%	11.4	4.02%	7.26%	
15.1	8.80%	2.50%	2.57%	15	14	13.11%	2.75%	2.82%	13	4.17%	10.86%	3.12%	17	11.23%	0.91%	1.93%	12	6.64%	3.45%	3.86%	12	3.96%	5.47%	
15.2	8.88%	2.51%	2.57%	15.3	10.5	9.72%	7.00%	2.84%	13.1	7.73%	8.62%	3.14%	18	12.13%	5.30%	1.94%	12.3	6.86%	4.72%	3.75%	13	4.62%	7.28%	
15.3	8.96%	2.53%	2.58%	16	15.33	14.40%	3.57%	2.88%	13.3	7.72%	8.64%	3.18%	19	13.02%	5.46%	1.94%	13	7.37%	3.48%	3.81%	14	4.37%	0.96%	
16	10.06%	2.14%	4.48%	16.2		13.66%	5.88%	2.89%	14	7.00%	10.40%	6.49%					13.3	7.60%	4.95%	3.80%	14.2	4.85%	7.31%	
16.2	9.68%	2.66%	2.61%	16.3	11	10.20%	1.72%	2.89%	14.3	7.67%	8.71%	3.35%					14	8.11%	10.75%	4.22%	15	5.09%	1.02%	
17	10.31%	2.86%	3.36%	17	16	15.05%	2.00%	2.93%	15	4.92%	2.67%	3.47%					15	8.85%	2.00%	3.90%	15.2	5.15%	7.32%	
17.1	10.40%	2.78%	2.64%	17.1		14.09%	5.85%	2.94%	16	7.58%	8.84%	3.65%					16	9.59%	5.57%	3.95%	16	4.71%	7.34%	
17.2	10.48%	2.80%	2.64%	17.3	12	11.17%	6.63%	2.95%	17	7.53%	8.91%	3.82%					17	10.33%	5.80%	4.01%	16.4	5.51%	7.34%	
18	11.02%	3.46%	1.62%	18	17	16.02%	1.99%	2.99%													17	5.73%	1.14%	
18.1	11.19%	2.92%	2.67%	18.3	13	12.14%	1.94%	3.01%													17.4	5.80%	7.36%	
18.2	11.27%	2.94%	2.67%	19		15.00%	5.79%	3.05%													18	6.05%	27.09%	
18.3	11.35%	2.95%	2.67%	19.3	14	13.11%	1.18%	3.07%													19	5.94%	1.52%	
19	11.91%	3.05%	2.70%	20.3		15.63%	5.75%	3.12%													20	6.58%	7.40%	
20	12.71%	3.19%	2.73%	21		15.96%	5.73%	3.16%													21	8.09%	1.13%	
20.1	12.79%	3.21%	2.73%																		22	8.59%	7.44%	
21	13.51%	3.33%	2.76%																		23	7.34%	7.45%	
																					24	7.77%	7.47%	
																					25	8.07%	7.49%	

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information

Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
5 OF 32

D16S539				D18S51				D2S1338				CSF1PO				Penta D				D18S51 continued			
Allele	BS	FS	HBS	Allele	BS	FS	HBS	Panel	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS
4	2.54%	2.00%		7	4.05%	1.64%	1.55%	10	5.61%	10.21%	0.56%	5	2.40%			2.2	0.81%			18.1	11.95%	2.48%	1.18%
5	3.26%	2.15%		8	4.76%	1.72%	1.52%	11	6.09%	10.22%	0.66%	6	3.27%			3.2	1.05%			18.2	12.02%	2.48%	1.18%
6	3.98%	2.30%		9	5.48%	1.79%	1.49%	12	6.57%	10.23%	0.75%	7	4.14%			5	1.48%			19	12.59%	1.83%	0.68%
7	4.70%	2.45%		9.2	5.62%	1.81%	1.48%	13	7.04%	10.24%	0.85%	7.3	4.40%			6	1.71%			19.2	12.74%	2.56%	1.14%
8	5.42%	2.60%		10	6.19%	1.87%	1.45%	14	7.52%	10.25%	0.95%	8	5.00%			6.4	1.81%			20	13.30%	4.82%	1.11%
8.3	5.64%	2.64%		10.2	6.33%	1.88%	1.45%	15	8.00%	10.26%	1.05%	8.1	5.09%			7	1.95%			20.2	13.45%	2.63%	1.11%
9	6.15%	1.06%		11	6.90%	1.94%	1.42%	16	8.66%	0.73%	1.14%	8.3	5.27%			7.4	2.05%			21	14.02%	2.69%	1.08%
9.3	6.36%	2.79%		11.2	7.04%	1.96%	1.41%	17	7.30%	20.48%	1.24%	9	5.87%			8	2.65%			21.1	14.09%	2.70%	1.08%
10	6.87%	1.97%		12	7.61%	2.69%	2.30%	18	8.73%	10.29%	1.34%	9.1	5.96%			8.2	2.24%			21.2	14.16%	2.71%	1.07%
10.3	7.08%	2.94%		12.2	7.75%	2.03%	1.38%	18.3	9.57%	10.30%	1.37%	10	6.74%			9	1.45%			22	14.73%	3.84%	1.05%
11	7.59%	5.00%		12.3	7.82%	2.04%	1.37%	19	9.24%	2.80%	1.43%	10.1	6.83%			9.2	2.48%			22.2	14.87%	2.78%	1.04%
11.3	7.80%	3.09%		13	8.32%	1.61%	1.17%	19.3	10.04%	10.31%	1.46%	10.2	6.91%			9.4	2.52%			23	15.44%	2.84%	1.01%
12	8.31%	3.90%		13.1	8.39%	2.10%	1.35%	20	10.50%	11.77%	1.53%	10.3	7.00%			10	2.53%			23.2	15.58%	2.86%	1.01%
12.1	8.38%	3.21%		13.2	8.47%	2.11%	1.34%	21	11.58%	3.98%	1.63%	11	7.61%			11	2.32%			24	16.15%	2.92%	0.98%
12.2	8.45%	3.23%		13.3	8.54%	2.12%	1.34%	22	9.85%	4.80%	1.72%	11.1	7.69%			11.1	2.93%			24.2	16.29%	2.93%	0.97%
13	9.03%	2.87%		14	9.03%	1.29%	0.96%	22.2	11.42%	10.34%	1.74%	11.3	7.87%			11.2	2.95%			25	16.86%	2.99%	0.95%
13.3	9.24%	3.39%		14.2	9.18%	2.18%	1.31%	23	11.69%	2.12%	1.82%	12	8.47%			12	2.43%			26	17.57%	3.07%	0.91%
14	9.75%	1.99%		15	9.75%	1.43%	0.81%	23.2	11.90%	10.35%	1.84%	12.1	8.56%			12.1	3.17%			27	18.29%	3.14%	0.88%
15	10.47%	3.64%		15.2	9.89%	2.26%	1.28%	23.3	11.95%	10.35%	1.85%	12.3	8.73%			12.2	3.19%						
16	11.19%	3.79%		15.3	9.96%	2.27%	1.27%	24	13.02%	4.45%	1.92%	13	9.34%			13	2.70%						
				16	10.46%	1.53%	0.77%	24.2	12.38%	10.36%	1.94%	13.3	9.60%			13.4	3.47%						
				16.1	10.53%	2.33%	1.25%	25	13.30%	16.07%	2.01%	14	10.21%			14	4.26%						
				16.2	10.60%	2.33%	1.24%	26	16.31%	10.38%	2.11%	15	11.07%			14.1	3.64%						
				16.3	10.67%	2.34%	1.24%	27	13.71%	10.39%	2.21%	16	11.94%			14.4	3.71%						
				17	11.17%	1.92%	0.85%	28	14.18%	10.40%	2.31%					15	4.38%						
				17.2	11.31%	2.41%	1.21%	29	14.66%	10.41%	2.40%					16	4.09%						
				17.3	11.38%	2.42%	1.21%									17	4.33%						
				18	11.88%	3.34%	0.69%									18	4.57%						

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

Manual Appendix for PowerPlex Fusion Stutter Information

Document ID: 57011

APPROVED BY
Nuclear DNA Technical Leader

PAGE
6 OF 32

TH01					vWA				D21S11				D7S820				D5S818				TPOX					
Allele	LUS	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS		
3	3	1.71%		1.32%	10	3.87%	6.71%		23.2	9.50%	3.03%		5	1.99%	3.58%	2.55%	5	1.27%	1.99%		4	0.93%				
4	4	2.27%		1.32%	11	4.83%	6.60%		23.3	9.54%	3.03%		5.2	2.16%	3.57%	2.54%	6	2.25%	2.13%		5	1.53%				
5	5	2.82%		1.31%	12	5.63%	6.48%		24	9.76%	3.05%		6	2.85%	3.55%	2.49%	7	3.24%	2.27%		6	2.13%				
5.3		4.13%		1.31%	13	6.76%	6.36%		24.2	9.83%	3.06%		6.3	3.10%	3.54%	2.48%	8	4.22%	2.41%		7	2.73%				
6	6	3.37%		1.07%	14	7.83%	11.80%		24.3	9.86%	3.06%		7	3.70%	3.52%	2.44%	9	5.20%	2.56%		7.1	2.79%				
6.1		4.16%		1.31%	15	8.46%	0.87%		25	10.09%	3.08%		7.1	3.79%	3.52%	2.44%	10	6.18%	2.10%		7.3	2.91%				
6.3	3	1.71%		1.31%	15.2	8.88%	6.11%		25.2	10.15%	3.09%		7.3	3.96%	3.52%	2.43%	10.1	6.28%	2.71%		8	2.78%				
7	7	3.92%		0.91%	16	6.41%	6.02%		25.3	10.18%	3.09%		8	4.56%	3.01%	2.26%	11	7.16%	2.84%		9	3.93%				
7.1		4.21%		1.30%	16.1	9.75%	6.01%		26	10.41%	3.11%		8.1	4.65%	3.49%	2.39%	11.1	7.26%	2.85%		10	4.53%				
7.3		4.22%		1.30%	17	9.42%	1.76%		26.2	10.48%	3.11%		8.2	4.73%	3.49%	2.38%	12	8.14%	2.62%		10.1	4.59%				
8	8	4.48%		1.70%	18	10.80%	8.98%		27	10.73%	3.14%		8.3	4.82%	3.49%	2.38%	12.1	8.24%	3.00%		10.3	4.71%				
8.3	5	2.82%		1.30%	18.2	11.78%	5.76%		27.1	10.77%	3.14%		9	5.42%	3.47%	1.45%	12.3	8.43%	3.03%		11	5.38%				
9	9	5.03%		0.88%	18.3	11.88%	5.75%		27.2	10.80%	3.14%		9.1	5.50%	3.46%	2.34%	13	9.12%	3.25%		12	7.12%				
9.1		4.30%	1.29%	19	12.19%	1.89%	28	7.85%	6.31%	9.2	5.59%	3.46%	2.33%	14	10.10%	3.27%	13	6.33%								
9.3	6	3.37%	1.51%	20	12.23%	1.76%	28.1	11.09%	3.17%	9.3	5.67%	3.46%	2.33%	15	11.08%	5.71%	14	6.93%								
10	10	5.58%	1.29%	21	11.17%	5.44%	28.2	11.12%	3.17%	10	6.28%	4.37%	2.08%	16	12.06%	3.55%	15	7.53%								
10.3	6	3.37%	1.29%	22	15.45%	5.32%	28.3	11.16%	3.17%	10.1	6.36%	3.44%	2.29%	17	13.04%	3.69%	16	8.13%								
11	11	6.13%	1.29%	23	16.41%	5.21%	29	10.19%	2.89%	10.3	6.53%	3.43%	2.28%	18	14.02%	3.84%										
13.3	8	4.48%	1.28%	24	17.38%	5.09%	29.1	11.42%	3.20%	11	7.13%	3.51%	1.75%	19	15.01%	3.98%										
				25	18.34%	4.97%	29.2	11.45%	3.20%	11.1	7.22%	3.41%	2.24%													
							29.3	11.48%	3.20%	11.2	7.30%	3.41%	2.23%													
							30	13.62%	2.41%	11.3	7.39%	3.40%	2.23%													
							30.1	11.74%	3.22%	12	7.99%	2.20%	2.04%													
							30.2	11.23%	2.47%	12.1	8.08%	3.38%	2.19%													
							30.3	11.80%	3.23%	12.3	8.25%	3.38%	2.18%													
							31	12.03%	3.25%	13	8.85%	3.36%	2.14%													
							31.1	12.06%	3.25%	13.1	8.93%	3.35%	2.14%													
							31.2	11.87%	2.62%	14	9.70%	3.33%	2.09%													
							31.3	12.13%	3.26%	14.1	9.79%	3.32%	2.09%													
							32	11.47%	3.28%	15	10.56%	3.30%	2.04%													
							32.1	12.39%	3.28%	16	11.42%	3.27%	1.99%													

© NYC OFFICE OF CHIEF MEDICAL EXAMINER

Qualtrax template 040621

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information

Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
7 OF 32


D8S1179				D12S391				D19S433				FGA				D22S1045				FGA continued			
Allele	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS	Allele	BS	FS	HBS
6	7.00%	2.87%		13	5.59%	3.40%		5.2	4.09%	16.66%		12.2	5.32%	4.79%		6	0.61%	9.37%		25	10.50%	7.63%	
7	7.50%	2.89%		14	6.74%	3.49%		6	4.68%	16.00%		13	5.69%	5.03%		7	2.08%	10.10%		25.1	11.24%	8.76%	
8	5.10%	2.91%		15	7.89%	3.58%		6.2	4.83%	15.83%		14	6.15%	5.34%		8	3.56%	10.83%		25.2	11.28%	8.79%	
9	8.51%	2.93%		16	9.05%	3.67%		7	5.43%	15.16%		15	6.60%	5.65%		9	5.03%	11.55%		25.3	11.33%	8.82%	
10	7.44%	2.15%		17	10.20%	2.00%		8	6.18%	14.33%		16	7.06%	5.96%		10	6.51%	12.28%		26	9.23%	9.03%	
11	11.98%	2.96%		17.1	10.32%	3.77%		8.2	6.33%	14.16%		16.1	7.11%	5.99%		11	7.98%	10.02%		26.1	11.69%	9.07%	
12	10.54%	4.18%		17.3	10.55%	3.79%		9	6.92%	13.49%		16.2	7.15%	6.02%		12	9.45%	16.28%		26.2	11.74%	9.10%	
12.3	10.16%	2.98%		18	11.35%	4.84%		9.2	7.07%	13.33%		17	7.52%	6.26%		13	10.93%	14.47%		26.3	11.79%	9.13%	
13	9.94%	3.32%		18.1	11.47%	3.86%		10	7.67%	12.66%		17.2	7.61%	6.33%		14	12.40%	7.11%		27	12.11%	9.34%	
13.2	10.61%	3.00%		18.3	11.70%	3.88%		10.2	7.82%	12.49%		18	7.98%	6.57%		15	13.88%	18.23%		27.2	12.20%	9.40%	
13.3	10.66%	3.00%		19	12.51%	3.94%		11	4.13%	11.82%		18.2	8.07%	6.63%		16	15.35%	15.12%		27.3	12.24%	9.43%	
14	10.56%	3.00%		19.1	12.62%	3.95%		11.1	8.49%	11.74%		19	6.62%	6.88%		17	16.82%	17.87%		28	12.57%	9.65%	
14.1	11.06%	3.02%		19.3	12.85%	3.97%		11.2	8.56%	11.66%		19.1	8.48%	6.91%		18	18.30%	18.12%		28.2	12.66%	9.71%	
15	10.44%	2.45%		20	13.66%	2.70%		12	10.62%	10.99%		19.2	8.53%	6.94%		19	19.77%	18.85%		28.3	12.70%	9.74%	
15.1	11.56%	3.03%		20.1	13.77%	4.04%		12.1	9.24%	10.91%		19.3	8.58%	6.97%		20	21.25%	19.58%		29	13.02%	9.96%	
15.3	11.66%	3.04%		20.3	14.00%	4.05%		12.2	9.31%	10.82%		20	8.05%	6.03%						29.1	13.07%	9.99%	
16	13.53%	2.81%		21	14.81%	1.38%		13	8.78%	23.50%		20.1	8.94%	7.22%						29.2	13.12%	10.02%	
17	12.47%	3.78%		21.3	15.16%	4.14%		13.1	9.98%	10.07%		20.2	7.70%	7.25%						30	13.48%	10.27%	
18	13.91%	1.35%		22	15.96%	2.66%		13.2	10.06%	9.99%		20.3	9.03%	7.28%						30.2	13.57%	10.33%	
19	13.51%	3.10%		22.3	16.31%	4.23%		14	9.47%	4.62%		21	8.87%	7.50%						31	13.94%	10.57%	
20	14.01%	3.12%		23	17.12%	1.70%		14.1	10.73%	9.24%		21.1	9.40%	7.53%						31.2	14.03%	10.64%	
				24	18.27%	10.35%		14.2	13.80%	6.21%		21.2	9.45%	7.56%						32	14.40%	10.88%	
				24.3	18.61%	4.41%		14.3	10.88%	9.07%		21.3	9.49%	7.59%						32.2	14.49%	10.94%	
				25	19.42%	4.47%		15	10.86%	1.85%		22	9.87%	3.11%						33.2	14.95%	11.25%	
				26	20.57%	4.56%		15.2	11.50%	3.96%		22.1	9.86%	7.83%						34.2	15.41%	11.56%	
				27	21.73%	4.65%		16	13.54%	3.94%		22.2	10.87%	7.87%						41.2	18.62%	13.71%	
								16.1	12.22%	7.57%		22.3	9.95%	7.90%						42.2	19.08%	14.02%	
								16.2	9.43%	3.15%		23	8.40%	7.26%						43.2	19.54%	14.33%	
								17	12.89%	6.82%		23.1	10.32%	8.14%						44.2	19.99%	14.64%	
								17.2	10.06%	6.65%		23.2	10.36%	8.17%						45.2	20.45%	14.95%	
								18	13.64%	5.98%		23.3	10.41%	8.20%						46.2	20.91%	15.25%	
								18.1	13.71%	5.90%		24	12.86%	6.65%						47.2	21.37%	15.56%	
								18.2	13.79%	5.82%		24.1	10.78%	8.45%						48.2	21.83%	15.87%	
								19	14.38%	5.15%		24.2	10.82%	8.48%						49.2	22.29%	16.18%	
								19.2	14.53%	4.98%		24.3	10.87%	8.51%						50.2	22.75%	16.48%	
																				51.2	23.21%	16.79%	

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information		
Status: Published		Document ID: 57011
DATE EFFECTIVE 06/23/2025	APPROVED BY Nuclear DNA Technical Leader	PAGE 8 OF 32

Table 5. Stutter ratio (SR) data collected for each location, allele, and stutter type. GeneMarker® filters were determined using a value two standard deviations above either the average value, the allele regression line, or the LUS regression line. Graphs show the stutter ratio on the y-axis and allele number of the parent allele or LUS value on the x-axis.



Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information



Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
9 OF 32



Allele	LUS Allele
11	11
12	11.5
13	12
14	13.5
14.3	9
15	14
15.3	10.5
16	15.33
16.3	11
17	16
17.3	12
18	17
18.3	13
19.3	14

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information

Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025


APPROVED BY
Nuclear DNA Technical Leader

PAGE
10 OF 32

(See LUS for BS)


D1S1656 FS

$$y = -0.0003x + 0.0182$$



D1S1656 HBS


$$y = 0.0006x + 0.0046$$



D1S1656 DBS

$$y = 0.0003x + 0.0031$$

No GeneMarker filter but modeled by STRmix



Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information





Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
11 OF 32



Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

© NYC OFFICE OF CHIEF MEDICAL EXAMINER

Qualtrax template 040621

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information




Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
12 OF 32



D10S1248 DBS

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information




Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
13 OF 32



D13S317 DBS

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information


Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025


APPROVED BY
Nuclear DNA Technical Leader

PAGE
14 OF 32



Penta E HBS

No GeneMarker filter
Not modeled by STRmix




Penta E DBS

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.


FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information		
Status: Published	Document ID: 57011	
DATE EFFECTIVE 06/23/2025	APPROVED BY Nuclear DNA Technical Leader	PAGE 15 OF 32



D16S539 HBS

No GeneMarker filter
Not modeled by STRmix



D16S539 DBS

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information





Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
16 OF 32



Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information





Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
17 OF 32




Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

© NYC OFFICE OF CHIEF MEDICAL EXAMINER

Qualtrax template 040621

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information		
Status: Published		Document ID: 57011
DATE EFFECTIVE 06/23/2025	APPROVED BY Nuclear DNA Technical Leader	PAGE 18 OF 32



No GeneMarker filter
Not modeled by STRmix

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information


Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
19 OF 32



Penta D FS

No GeneMarker filter
Not modeled by STRmix

Penta D HBS

Penta D DBS

No GeneMarker filter
Not modeled by STRmix

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

© NYC OFFICE OF CHIEF MEDICAL EXAMINER

Qualtrax template 040621

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information



Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
20 OF 32



Allele	LUS Allele
6	6
7	7
8	8
9	9
9.3	6
10	10

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information

Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
21 OF 32


TH01 FS

(See LUS for BS)

No GeneMarker filter
Not modeled by STRmix

TH01 HBS

TH01 DBS



No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information


Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025


APPROVED BY
Nuclear DNA Technical Leader

PAGE
22 OF 32



vWAHBS

No GeneMarker filter
Not modeled by STRmix





vWADBS

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.


FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information		
Status: Published	Document ID: 57011	
DATE EFFECTIVE 06/23/2025	APPROVED BY Nuclear DNA Technical Leader	PAGE 23 OF 32



D21S11 HBS

No GeneMarker filter
Not modeled by STRmix



Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information




Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025

APPROVED BY
Nuclear DNA Technical Leader

PAGE
24 OF 32



D7S820 DBS

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information


Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025


APPROVED BY
Nuclear DNA Technical Leader

PAGE
25 OF 32



D5S818 HBS

No GeneMarker filter
Not modeled by STRmix




D5S818 DBS

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information		
Status: Published		Document ID: 57011
DATE EFFECTIVE 06/23/2025	APPROVED BY Nuclear DNA Technical Leader	PAGE 26 OF 32



TPOX FS

No GeneMarker filter
Not modeled by STRmix

TPOX HBS

No GeneMarker filter
Not modeled by STRmix

TPOX DBS

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information



Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025


APPROVED BY
Nuclear DNA Technical Leader

PAGE
27 OF 32



D8S1179 HBS

No GeneMarker filter
Not modeled by STRmix



Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information



Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025


APPROVED BY
Nuclear DNA Technical Leader

PAGE
28 OF 32



D12S391 HBS


No GeneMarker filter
Not modeled by STRmix



Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.


FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information		
Status: Published	Document ID: 57011	
DATE EFFECTIVE 06/23/2025	APPROVED BY Nuclear DNA Technical Leader	PAGE 29 OF 32



D19S433 HBS

No GeneMarker filter
Not modeled by STRmix



D19S433 DBS

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information



Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025


APPROVED BY
Nuclear DNA Technical Leader

PAGE
30 OF 32



FGA HBS

No GeneMarker filter
Not modeled by STRmix



Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information


Status: Published

Document ID: 57011

DATE EFFECTIVE
06/23/2025


APPROVED BY
Nuclear DNA Technical Leader

PAGE
31 OF 32



D22S1045 HBS

No GeneMarker filter
Not modeled by STRmix



D22S1045 DBS

No GeneMarker filter
Not modeled by STRmix

Controlled versions of Department of Forensic Biology Manuals only exist in the Forensic Biology Qualtrax software. All printed versions are non-controlled copies.

© NYC OFFICE OF CHIEF MEDICAL EXAMINER

Qualtrax template 040621

FORENSIC BIOLOGY PROTOCOLS FOR FORENSIC STR ANALYSIS

Manual Appendix for PowerPlex Fusion Stutter Information		
Status: Published		Document ID: 57011
DATE EFFECTIVE 06/23/2025	APPROVED BY Nuclear DNA Technical Leader	PAGE 32 OF 32

References

- 1 Office of the Chief Medical Examiner- NYC PowerPlex® Fusion System Amplification Kit on the Applied Biosystems 3130xl Genetic Analyzer using GeneMarker® HID Analysis Software Validation Report, August 29, 2016.
- 2 NYC OCME Stutter Study for GeneMarker® HID 3.0.0 and STRmix™ Version 2.7-; PowerPlex® Fusion Data run on 3500xl Genetic Analyzers
- 3 Estimation of STRmix™ parameters for OCME New York Laboratory
- 4 Internal Validation of STRmix™ V2.4 for Fusion NYC OCME